



ISO 9001 / ISO 13485 / ISO 14001 / OHSAS 18001 / CE Certified / WHO GMP Complaint Co / US FDA Establishment Regn No : 3005141524

ANAESTHESIA PRODUCTS



Improving patient care worldwide is our concern

Heat & Moisture Exchange Filter



- Tidal volume range (ml) : 70-600
 - Compressible volume (Dead space) (ml) : 22
 - Humidification efficiency (mg/l)
at Tidal Volume = 100 ml:33.0 at Tidal Volume = 250 ml:31.0
at Tidal Volume = 500 ml:27.5 at Tidal Volume = 1000 ml
 - Bacterial Filtration (%) : 99.98
 - Viral Filtration (%) : 99.80
 - Resistance (cm H₂O) at flow :
0.15 l/s: 0.6 0.25 l/s: 1.0 0.50 l/s: 2.1 1.0 l/s 1.5 l/s
 - Weight (g) : 16.5
- Packing** : Box of 100 pcs.
Size : Adult & Paediatric.

LIFELINE SYSTEMS PVT. LTD.

Regd. Office : A.D. House, J-161/1, Gautam Nagar, New Delhi-110049. INDIA.
Sales Office : F-63, Okhla Industrial Area Phase-1, New Delhi-110020. INDIA.
Phone : +91-11-41640924 - 6, 40523028
Fax : +91-11-40523027, 26810652
Email : info@lifelinedelhi.com
URL : www.lifelinedelhi.com

Branch Offices : 111/506, Navjivan Society, Lamington Road, Mumbai-400008. India.
Phone : +91-22-23096796
Sundaram Building, 1-D, 1st Floor, 46 Rafi Ahmed Kidwai Road,
Kolkatta-700016. India. Phone : +91-33-22299932